

Titel **MASPAT kit release**

K 1360

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL
CL

MASPAT kit
Monoclonal Antibody Solid Phase
Platelet Antibody Test

  96
LOT 8000460484
REF K1360

  

 2-8 °C
 2026-01-10


(01)1872068203
0025(10)800046
0484(17)260110

Component	Requirement	Result
Positive control Lot: 8000460309	Positive reaction with: platelets Titration with: platelets $\geq 1:4$	Conform requirements Conform requirements
Negative control Lot: 8000460310	Negative reaction with: platelets	Conform requirements
Anti-IgG Lot: 8000460227	Titration with: platelets $\geq 1:4$	Conform requirements
Microplates Lot: 8000460324	Specificity: Platelets x positive control positive Platelets x negative control negative Coating tested platelets for reactivity and homogeneity	Conform requirements Conform requirements Conform requirements
LISS Lot: 8000460308	Conductivity 3.4 – 4.8 mS cm^{-1} pH 6.3 – 6.9	Conform requirements Conform requirements
Growth test	no growth	Conform

All the sera used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with the current EU guidelines to GMP and found to be non reactive

Date : 05 April 2024 Approved by QA/QC : 

VERTROUWELIJK