



Doc. No. REA-S-0080
Date 24 Feb 2023

Revision 03
Page 1 of 1
Document Owner BBR

Title Pelikloon anti-D mix (IgG-IgM) monoclonal release

K1157
Clone numbers TH28/MS26



Analysis	Requirement			Result
Specificity	Negative with Rh(D)negative, r'r and r''r cells			Conform requirements
Avidity		Dd	Dweak	
	Tube method	≥ 2+	≥ W+	Conform requirements
	Microplate method	≥ 2+	≥ W+	Conform requirements
	Cellbind Screen method	≥ 2+	≥ 1+	Conform requirements
Titer		Dd	Dkl.VI	
	Tube method	≥ 1:32	≥ 1:4 (IAT)	Conform requirements
	Microplate method	≥ 1:32	n.a.	Conform requirements
Growthtest	no growth			Conform
External release	released by Notified Body			Released Y/N

Date: 01 May 2024

Approved by

QA/QC:

CONFIDENTIAL