



Doc. nr. REA-S004-01-17-0529

Versie 2.0


Datum 06 May 2021 13:54:34 (GMT+02:00)

Pagina 1 van 2

Document eigenaar BBR

Titel anti-s release

K 1343



Analysis	Requirement	Result
Specificity	Negative with A, B and O cells negative for the s antigen	Conform requirements
Avidity	Tube method	
	ss $\geq 3+$	Conform requirements
	Ss $\geq 3+$	Conform requirements
Titer	Tube method	
	Ss $\geq 1:8$	Conform requirements
Growthtest	no growth	Conform

All the sera used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be nonreactive.

Date

: 26 April 2024

Approved by QA/QC

:

VERTROUWELIJK

Use copy according to REA-SOP-0002.
Printed on: 22 Apr 2024 at: 07:44:27 am ; Printed by B.SMIT